

实验室安全检查通报

东北林业大学 实验室管理与条件保障处

检查内容：实验室安全检查

检查时间：2026年4月28日

检查范围：动保学院教学、科研实验室

检查人员：校级实验室安全督查员

动保学院：

4月28日，由学校实验室安全督查员吴德东、黄海娇及相关人员对动保学院的科研、教学实验室进行了检查，现场检查中发现主要存在仪器设备管理、用电基础管理等方面的安全隐患，请学院务必对此项进行重视，确保实验室安全稳定运行。在现场检查中发现存在以下具体问题，见表1。

表1 检查中发现的实验室安全管理中存在的问题

| 检查日期 | 实验室 房间号 | 检查问题记录 |
|------------|------------|---|
| 2026年4月28日 | 动保楼001 | 1、气瓶未进行合理固定；未配备气瓶状态卡； 2、烘箱高温区域周边密集放置化学试剂瓶，试剂瓶紧邻热源，存在受热炸裂、试剂泄漏及火灾安全隐患； 3、窗台侧面违规放置插线板，易受潮进水引发漏电、触电风险；未核算配电负荷，未确保接线板容量与用电设备总功率相匹配，存在过载起火隐患； 4、超净工作台上方存放纸箱，阻挡空气过滤，影响设备使用； 5、电泳区域没有警示标识。 |
| 2026年4月28日 | 动保楼009 | 1、制毒重点设备未按要求进行管理，未严格执行设备使用登记、台账记录。 |
| 2026年4月28日 | 动保楼702 | 1、制毒重点设备使用登记台账记录不完整、信息填写不规范； |

| | | |
|------------|--------|--|
| | | 2、实验废液未严格遵照规范进行分类收集，未按要求分装至专用废液储存桶。 |
| 2026年4月28日 | 动保楼716 | 1、制毒重点设备使用登记台账记录不完整、信息填写不规范； 2、实验室未配备合适的灭火设备。 |

请学院督促以上实验室立即进行整改，并撰写整改报告（整改报告模板见实验室管理与条件保障处网站<https://elab.nefu.edu.cn/info/1052/1633.htm>），于5月9日前提交实验室管理与条件保障处备案，其他实验室举一反三，排查问题，杜绝隐患。




办公地点：综合办公楼 604

办公电话：82190708

实验室管理与条件保障处

2026年4月28日

实验室安全隐患图

| 房间号 | 隐患图 |
|------|--|
| 001: |  <p>The images for room 001 show several safety concerns: a gas cylinder in a confined space, a machine with bottles nearby, a cluttered workbench with various equipment, a storage area with boxes, and a lab bench with various reagents and containers.</p> |
| 009: |  <p>The image for room 009 shows a white machine with a red fire extinguisher nearby, indicating a potential fire hazard.</p> |
| 702: |  <p>The images for room 702 show several safety concerns: handwritten lab notes, a bottle of liquid, and a cardboard box labeled '废液' (waste liquid), indicating improper storage of hazardous materials.</p> |

716:

