

实验室安全检查通报

东北林业大学 实验室管理与条件保障处

检查内容：实验室安全检查

检查时间：2025 年 7 月 3 日

检查范围：理学院教学科研实验室

检查人员：马洪斌、郭敏、李凯

理学院：

7月3日，实验室管理处副处长马洪斌及相关人员对理学院教学科研实验室进行安全检查，发现实验室在日常管理、基础用电等方面存在安全隐患。针对实验室确需使用接线板为加热设备供电的情况，明确要求严禁超负荷使用且使用时必须有人值守。临近假期，请学院高度关注实验室隐患整改工作，严格落实安全管理措施，加大隐患排查与整改力度，确保及时消除安全风险，保障实验室安全稳定运行。具体问题如下。

实验室存在的问题

检查日期	实验室房间号	检查问题记录
2025年7月3日	103	<ol style="list-style-type: none">实验室有饮食行为，实验台上有饮料瓶；未按管理要求，私自使用插排连接多台加热设备供电，（烘箱、电阻炉功率大、电流大，如使用接线板供电，易造成接线板过载，导致火灾事故），存在消防安全隐患。实验室存放大量纸箱易燃物品，遇到明火、高温或电火花等火源极易燃烧，可能引发火灾；实验室卫生环境差；实验室风险评估不全面，请实验室全面排查各类危险源，动态调整实验室风险等级，完善防控措施。

请学院督促实验室立即进行整改，并撰写整改报告（整改报告模板见实验室管理与条件保障处网站<https://elab.nefu.edu.cn/info/1052/1633.htm>），于7月9日前提交实验室管理与条件保障处备案，其他实验室举一反三，排查问题，杜绝隐患。

办公地点：综合办公楼 604

办公电话：82190708

实验室管理与条件保障处

2025年7月4日

实验室安全隐患图

实验室名称	隐患图
103	 <p>The image contains four photographs illustrating safety hazards in a laboratory. The top-left photo shows a cluttered workbench with a black bag, a green bottle, and various papers. The top-right photo shows a blue piece of equipment with tangled cables and scattered pipette tips. The bottom-left photo shows a white cabinet with tools and papers on top. The bottom-right photo shows a stack of cardboard boxes, including one labeled 'IKA' and another '磁力搅拌器' (magnetic stirrer).</p>